



465 Old Pelzer Road
 Piedmont, SC 29673
 (864) 859-1471

CERTIFICATE OF ANALYSIS

GLYCOLIDE MONOMER

Ortec Product Number: 052-02-003
 Lot/Batch Number: 02220600076
 Date of Manufacture: 06/28/2022
 Recommended Retest Date: 09/28/2022 (Three (3) months from Date of Manufacture)

TEST	RESULT	UNIT	SPECIFICATION	TEST METHOD
Monomer Appearance	White, crystalline solid	N/A	White, crystalline solid	PLAB-test-061
Moisture Content	51	ppm	< 200 ppm	PLAB-test-010
Free Acid	0.41	meq/kg	≤ 1.00 meq/kg	PLAB-test-021
Melt Point	85	°C	82 - 86°C	PLAB-test-023
Purity	99.9	%	≥ 99.5% by area	PLAB-test-025
Solvent Content	419	ppm	< 2000 ppm	PLAB-test-024

Recommended Storage Conditions: < -10°C

Compliance Statement:

This lot of Glycolide was manufactured in accordance with applicable regulatory and ISO requirements and is in compliance with applicable Ortec procedures and specifications. This product meets the approved release specification in place at Ortec.

Prepared by: Nancy Huch Date Prepared: 07-11-2022
 Quality Representative

Reviewed by: Serry Helton Date Reviewed: 7-11-2022

Ortec, Inc. is not responsible for use of information contained herein as to the determination and use of product.
 Customer assumes responsibility for fitness of use and merchantability of products to be used therein.
 Copyright © by Ortec, Inc. All rights reserved.