



No. AP018

Certificate of Analysis



Product Name: Poly(lactide-co-glycolide) acid endcapped (75:25 LA:GA)

(M_n: 45,000-55,000 Da) (Lot#: 180601RAI-A)



GPC Analysis Method: Waters Breeze 2 system with 1 ml/min THF flow across three GPC columns. Detection via refractive index, calibrated against polystyrene standards.

Polymer	M _n (from GPC)	M _w (from GPC)	PDI
PLGA	50,977	91,595	1.80

• Structure of PLGA

