



No. AP178

Certificate of Analysis



Product Name: Poly(DL-Lactic-co-caprolactone) copolymer LA:CL 70:30 acid endcap (M_n 15,000-25,000 Da)(Lot#: 180207JSG-A)



GPC Analysis Method: Waters Breeze 2 system with 1 ml/min THF flow across three GPC columns. Detection via refractive index, calibrated against polystyrene standards.

Polymer	M_n (from GPC)	M_w (from GPC)	PDI
P(D,L)LaCL	17,395	22,697	1.30

- Structure of P(D,L)LaCL copolymers**

